

WARNING

for service.

with skin infections.

Never under any circumstances attempt

to open or inspect the internal

components or accessories of the

machine. If an inspection becomes

or a suitably competent technician.

necessary, please contact your supplier

Never use the unit in humid or poorly

damage to the electrical components of

• Never use the unit outdoors; exposure to

If the power cord was destroyed, please

ask a professional to service and replace

it to avoid hazards, or contact the dealer

• Do not use on pregnant women, on

people with pace-makers (any electrical

or rhythmic heart-regulating device) or

ventilated environments, such as wet

rooms, bathrooms, etc. to prevent

the elements can cause damage.

ΕN

WKB002

PRESSOTHERAPY& THERMOTHERAPY

Instruction Manual WE-2020-WKB002-1

PACKAGE CONTENTS

incorrect operation.

- These components are subject to change without prior notice.
- Appearance of product may differ from the image.

The company reserves the right to amend


the details of the instruments with no prior

notice. If any mistakes are found in these

instructions, please feel free to send us an

e-mail informing us so that we can correct

No.	PRODUCT		QTY.
1	RIGHT LEG		1
2	LEFT LEG		1
3	RIGHT ARM		1
4	LEFT ARM		1
5	RIGHT FOOT		1
6	LEFT FOOT		1
7	WAIST		1
8	CABLE BLANKET		1
9	TUBES		1
10	MAIN UNIT	-	1
11	FEEDER CABLE	-	1



EC-DECLARATION OF COMPLIANCE

This product meets the following requirements:

- 1. Conditions set out in the Low Voltage Directive: 2014/35/EU
- 2. Conditions set out in the Electromagnetic Compatibility Directive: 2014/30/EU

RECYCLING

Whenever any of our electrical products or their presentation Whenever any of our electrical packaging bear this symbol, it means that these products should not be treated as conventional household waste in Europe. To ensure proper treatment of this waste, dispose of it in accordance with local laws or as required for the disposal of electrical equipment. This initiative will help preserve save natural resources and improve standards of environmental protection in regard to the treatment of electrical waste.

WELCOME!

Thank you very much for purchase. Please read the instructions carefully before operating the instrument. You can also ask a professional for guidance to make sure you are using the instrument correctly. We will accept no responsibility for any accidents that occur due to

INSTRUCTIONS OF USE

- 1. Connect the power cord via the rear panel.
- 2. Connect the tubes properly to the rear panel. Follow the numerical instructions. Thinner cables from 1 to 5 are used for thermotherapy and thicker cables from 0 to 9 are used for pressotherapy.
- 3. Place the blanket with the cables on a flat surface so that the patient can lay down on it. Cover the blanket with disposable paper to ensure the best hygienic conditions.
- 4. Wrap the suit around the user's body. Remember to use a disposable garment underneath the pressotherapy suit.
- 5. Turn on the power button on the front panel and click on Weelko logo to select the desired menus.
- **6.** Select "pressotherapy" for pressotherapy treatments or "far infrared" for thermotherapy treatments.

pressotherapy

far infrared

mode

between four programs. It is

patient's needs.

again.

7. In the pressotherapy function, the

"MODE" menu allows you to choose

recommended to start with C and to select the other programs depending on

section to another. At the end of the

program, sections from 0 to 9 are inflated

• Mode B fills with air sections from 0 to 9.

Section 0 is deflated when section 2 is

section 3 is inflated and so on from one

sections from 0 to 9 are inflated at the

same time and the program starts again.

• Mode C fills with air sections from 0 to 9. First section 0 is inflated, then section 1, later section 2 and so on through all

sections. At the end, the program starts

• Mode D fills with air the three sections at

the same time: legs, abdomen and arms

again from section 0.

from bottom to top.

side to another. At the end of the program,

inflated. Section 1 is deflated when

• Mode A fills with air sections from 0 to 9.

level

If you don't want to use a concrete mode, the device allows to choose sections separately and to adjust the inflation intensity of each section. You just need to press the area you want to treat and adjust its intensity by using the buttons "+" and "-".

- 8. The "LEVEL" menu allows to adjust 3 inflation intensity levels.
- 9. The "TIME" menu allows to adjust the treatment time. The default duration is 30 minutes.
- **10.** In the thermotherapy function, the device is equipped with 5 treatment sections. Temperature can be adjusted separately in each section. The "TIME" menu is used to adjust the treatment time. The default duration is 30
- 11. Press "PLAY/PAUSE" to begin or pause the treatment.

minutes.

- at the rear part of the device should be connected and disconnected.
 - After a thermotherapy treatment, open the suit to remove stored heat during the treatment.

• Connect and disconnect the tubes

carefully to extend the product lifetime.

However, this device is supplied with a

all connected and disconnected at the

same time. Only the general power cord

• After use, clean the suit with a wet cloth.

blanket to store the cables so that they are

CAUTIONS

MAINTENANCE

- Be careful not to wrap the airbag too tight, to avoid hurting the users when the air is too strong.
- If the surface of the blanket is broken or aging, please replace it with a new blanket to prevent accidental skin burns or electric shocks.
- During the treatment, check with the patient if the temperature is appropriate.
- It is recommended for the patient to drink plenty of water before and after treatment to avoid dehydration and heatstroke.
- Use a cellulose protector underneath the pressotherapy suit to ensure the best hygienic conditions.
- Never leave the suit heating with nobody inside.

ACCESSORIES


RIGHT LEG
LEFT LEG
RIGHT ARM
LEFT ARM
WAIST
RIGHT FOOT
LEFT FOOT
TUBES


TECHNICAL SHEET

(DII)

play/pause

Weights and sizes	Packing size (cm)	66x45x40
	Gross weight (kg)	18,5
	Product size (cm)	34x38x15
Working characteristics	Working power	350 W
	Air pressure	0-0,4 kg/cm2
	Air flow	0,40 L/min.
	Rhythm of air pressure	0-15 s.
	Air channels	10 (20 bags)
	Temperature range	30-50°C
Electrical characteristics	Rated voltage	220V ~ 240V
	Frequency	50Hz/60Hz
	Rated power	350 W





Section 0 is deflated when section 1 is inflated. Section 1 is deflated when section 2 is inflated and so on from one at the same time and the program starts

© 2020 Weelko Visit our website at www.weelko.com for more information